Andor Technology Ltd.

7 Millennium Way, Springvale Business Park, Belfast, BT12 7AL, UK

Tel: +44 (0)28 9023 7126 Fax: +44 (0)28 9031 0792 andor.com





Object of the Declarations

Dragonfly 300, 500 & 600 Series Confocal Microscopy Spinning Disk Unit

These declarations of conformity are issued under the sole responsibility of Andor Technology Ltd. who manufacture this at the above address.

UKCA and UK(NI) Marks

As this is a Qualifying Northern Ireland Good that is self-declared, this product does not require a UKCA Mark, UK(NI) Mark or UK Declarations of Conformity.

EU Declaration of Conformity (EMC)

The object of the declaration described above is in conformity with the **EMC Directive 2014/30/EU** by means of conformity to the following harmonised standards:

- EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use EMC requirements Part 1: General requirements [Class A Group 1 Basic Immunity]
- EN 55011:2016 + A1:2020 Industrial, scientific and medical (ISM) radio-frequency equipment – Electromagnetic disturbance characteristics – Limits and methods of measurement [Class A Group 1]
- EN 61000-4-2:2009 Electromagnetic compatibility (EMC) Testing and measurement techniques Electrostatic discharge immunity test [Criterion A]
- EN 61000-4-3:2020 Electromagnetic compatibility (EMC) Testing and measurement techniques Radiated, radio-frequency, electromagnetic field immunity test [Criterion A]
- EN 61000-4-4:2012 Electromagnetic compatibility (EMC) Testing and measurement techniques Electrical fast transient/burst immunity test [Criterion A]
- EN 61000-4-5:2014+A1:2017 Electromagnetic compatibility (EMC) Testing and measurement techniques Surge immunity test [Criterion A]
- EN 61000-4-6:2014 Electromagnetic compatibility (EMC) Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields [Criterion A]
- EN 61000-4-11:2020 Electromagnetic compatibility (EMC) Testing and measurement techniques Voltage dips, short interruptions and voltage variations immunity tests [Voltage Dips: Criterion B/C] [5-second Short Interruption: Criterion C]



EU Declaration of Conformity (LVD)

The object of the declaration described above is in conformity with the **Low Voltage Directive 2014/35/EU** by means of conformity to the following harmonised standards:

- EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements (identical to IEC 61010-1 3rd Edition)
- EN 60825-1:2014 Safety of laser products Part 1: Equipment classification and requirements

EU Declaration of Conformity (Machinery Directive)

The object of the declaration described above is in conformity with the **Machinery Directive 2006/42/EC** as well as by means of conformity to the following harmonised standard:

• EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements (identical to IEC 61010-1 3rd Edition)

The Product Compliance Advisor, Timothy Davis, at the address above is authorised to compile the technical file.

EU/UK REACH Statement

In relation to **REACH Regulation (EC) No. 1907/2006**, this product contains SVHCs in a concentration above 0.1 % weight by weight (w/w) in an 'article'. Please see <u>Andor's REACH Statement</u> for details.

We continue to work with our suppliers on an on-going basis to monitor our supply chain for the relevant levels of SVHCs in any of our articles.

EU Declaration of Conformity (RED)

The object of the declaration described above is in conformity with the **Radio Equipment Directive (RED) 2014/53/EU** by means of conformity to the following harmonised standard:

• EN 300 330-2 V2.1.1 Electromagnetic compatibility and Radio spectrum Matters (ERM); Short Range Devices (SRD); Radio equipment in the frequency range 9 kHz to 25 MHz and inductive loop systems in the frequency range 9 kHz to 30 MHz; Part 2: Harmonized EN covering the essential requirements of article 3.2 of the R&TTE Directive

EU Declaration of Conformity (RoHS)

The object of the declaration described above is in conformity with **Directive 2011/65/EU** on the Restriction of the use of certain Hazardous Substances in electrical and electronic equipment as **amended by Directive (EU) 2015/863**.

Additional EMC Standards

This product also complies with the following:

- FCC Part 15 Subparts A and B Code of Federal Regulations
 Title 47: Telecommunications Part 15: Radiofrequency Devices technical requirements
 [Class A up to 1 GHz Not tested beyond this]
 - No FCC Mark as exempted under 47 CFR §15.103(c) "A digital device used exclusively as industrial, commercial, or medical test equipment."

Additional Safety Standards

This product has been found compliant under the IEC CB Scheme with IEC 61010-1 3rd Edition and all national deviations listed for Canada, Japan, Switzerland, UK and USA, including the following:

- UL 61010-1 Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements (3rd Edition 2012)
- CAN/ CSA-C22.2 No. 61010-1-12 Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements (3rd Edition)
- Code of Federal Regulations Title 21: Food and Drugs Part 1040: Performance Standards for Light-Emitting Products §1040.10 Laser products and 1040.11 Specificpurpose Laser Products (FDA CDRH)

CE Mark

CE Mark first applied to 300 & 500 Series in the year 2016, and to the 600 Series in the year 2022.

Claire Greenwood

Director of Engineering

Clave Greenwood.

11th October 2022

Belfast U.K.